




CVE-2013-2808

Published on: 10/05/2013 12:00:00 AM UTC

Last Modified on: 03/23/2021 11:28:42 PM UTC

CVE-2013-2808

[Source: Mitre](#)[Source: NIST](#)[CVE.ORG](#)[Print: PDF](#) 

Certain versions of [Xperconnect](#) from [Philips](#) contain the following vulnerability:

Heap-based buffer overflow in Xper in Philips Xper Information Management Physiomonitring 5 components, Xper Information Management Vascular Monitoring 5 components, and Xper Information Management servers and workstations for Flex Cardio products before XperConnect 1.5.4.053 SP2 allows remote attackers to execute arbitrary code via a crafted HTTP request to the Connect broker on TCP port 6000.

CVE-2013-2808 has been assigned by [@ics-cert@hq.dhs.gov](#) to track the vulnerability

CVSS2 Score: **9.3 - HIGH**

Access Vector	Access Complexity	Authentication
NETWORK	MEDIUM	NONE
Confidentiality Impact	Integrity Impact	Availability Impact
COMPLETE	COMPLETE	COMPLETE

CVE References

Description	Tags	Link
Philips Xper Buffer Overflow Vulnerability ICS-CERT	US Government Resource ics-cert.us-cert.gov text/html	MISC ics-cert.us-cert.gov/advisories/ICSA-13-277-01

By selecting these links, you may be leaving CVEreport webspace. We have provided these links to other websites because they may have information that would be of interest to you. No inferences should be drawn on account of other sites being referenced, or not, from this page. There may be other websites that are more appropriate for your purpose. CVEreport does not necessarily endorse the views expressed, or concur with the facts presented on these sites. Further, CVEreport does not endorse any commercial products that may be mentioned on these sites. Please address comments about any linked pages to comment@cve.report.

There are currently no QIDs associated with this CVE

Known Affected Configurations (CPE V2.3)						
Type	Vendor	Product	Version	Update	Edition	Language
Application	Philips	Xperconnect	All	All	All	All
Hardware 	Philips	Xper Flex Cardio	-	All	All	All
Hardware 	Philips	Xper Flex Cardio	-	All	All	All
Application	Philips	Xper Information Management Physiomonitoring 5	-	All	All	All
Application	Philips	Xper Information Management Physiomonitoring 5	-	All	All	All
Application	Philips	Xper Information Management Vascular Monitoring 5	-	All	All	All
Application	Philips	Xper Information Management Vascular Monitoring 5	-	All	All	All
cpe:2.3:a:philips:xperconnect:*:*:*:*:*:						
cpe:2.3:h:philips:xper_flex_cardio:*:*:*:*:*:						
cpe:2.3:h:philips:xper_flex_cardio:*:*:*:*:*:						
cpe:2.3:a:philips:xper_information_management_physiomonitoring_5:*:*:*:*:*:						
cpe:2.3:a:philips:xper_information_management_physiomonitoring_5:*:*:*:*:*:						
cpe:2.3:a:philips:xper_information_management_vascular_monitoring_5:*:*:*:*:*:						
cpe:2.3:a:philips:xper_information_management_vascular_monitoring_5:*:*:*:*:*:						

No vendor comments have been submitted for this CVE