




CVE-2018-16496

Published on: 05/26/2021 12:00:00 AM UTC

Last Modified on: 06/07/2021 03:09:00 PM UTC

CVE-2018-16496

[Source: Mitre](#)[Source: NIST](#)[CVE.ORG](#)[Print: PDF](#) 

Certain versions of [Versa Director](#) from [Versa-networks](#) contain the following vulnerability:

In Versa Director, the un-authentication request found.

CVE-2018-16496 has been assigned by [h1](#) support@hackerone.com to track the vulnerability - currently rated as **MEDIUM** severity.

CVSS3 Score: **5.3 - MEDIUM**

Attack Vector	Attack Complexity	Privileges Required	User Interaction
NETWORK	LOW	NONE	NONE
Scope	Confidentiality Impact	Integrity Impact	Availability Impact
UNCHANGED	NONE	LOW	NONE

CVSS2 Score: **5 - MEDIUM**

Access Vector	Access Complexity	Authentication
NETWORK	LOW	NONE
Confidentiality Impact	Integrity Impact	Availability Impact
NONE	PARTIAL	NONE

CVE References

Description	Tags	Link
HackerOne	hackerone.com text/html	h1 MISC hackerone.com/reports/1168193

By selecting these links, you may be leaving CVEreport webspace. We have provided these links to other websites because they may have information that would be of interest to you. No inferences should be drawn on account of other sites being referenced, or not, from this page. There may be other websites that

are more appropriate for your purpose. CVEreport does not necessarily endorse the views expressed, or concur with the facts presented on these sites. Further, CVEreport does not endorse any commercial products that may be mentioned on these sites. Please address comments about any linked pages to comment@cve.report.

There are currently no QIDs associated with this CVE

Known Affected Configurations (CPE V2.3)

Type	Vendor	Product	Version	Update	Edition	Language
Application	Versa-networks	Versa Director	-	All	All	All
cpe:2.3:a:versa-networks:versa_director:-:*:*:*:*:*:						

No vendor comments have been submitted for this CVE

[← Previous ID](#)

[Next ID →](#)