



CVE-2018-7510


Published on: 06/06/2018 12:00:00 AM UTC

Last Modified on: 03/23/2021 11:24:39 PM UTC

CVE-2018-7510

[Source: Mitre](#)[Source: NIST](#)[CVE.ORG](#)[Print: PDF](#)

CVSS:30/AV:N/AC:L/PR:N/UI:N/S:U/C:H/I:H/A:H



Certain versions of [Scroll Medical Air Systems](#) from [Beaconmedaes](#) contain the following vulnerability:

In the web application in BeaconMedaes TotalAlert Scroll Medical Air Systems running software versions prior to 4107600010.23, passwords are presented in plaintext in a file that is accessible without authentication.

CVE-2018-7510 has been assigned by [ics-cert@hq.dhs.gov](#) to track the vulnerability - currently rated as **CRITICAL** severity.

Affected Vendor/Software: [ICS-CERT](#) - [BeaconMedaes TotalAlert Scroll Medical Air Systems web application](#) version **All versions prior to version 4107600010.23**

CVSS3 Score: **9.8 - CRITICAL**

Attack Vector	Attack Complexity	Privileges Required	User Interaction
NETWORK	LOW	NONE	NONE
Scope	Confidentiality Impact	Integrity Impact	Availability Impact
UNCHANGED	HIGH	HIGH	HIGH

CVSS2 Score: **5 - MEDIUM**

Access Vector	Access Complexity	Authentication
NETWORK	LOW	NONE
Confidentiality Impact	Integrity Impact	Availability Impact
PARTIAL	NONE	NONE

CVE References

Description	Tags	Link
BeaconMedaes TotalAlert Scroll Medical Air Systems CISA	Third Party Advisory US Government Resource	MISC ics-cert.us-cert.gov/advisories/ICSMA-18-144-01

comment@cve.report

There are currently no QIDs associated with this CVE

Known Affected Configurations (CPE V2.3)

Type	Vendor	Product	Version	Update	Edition	Language
Hardware 🇹🇼 ↗️	Beaconmedaes	Scroll Medical Air Systems	-	All	All	All
Hardware 🇹🇼 ↗️	Beaconmedaes	Scroll Medical Air Systems	-	All	All	All
Operating System	Beaconmedaes	Scroll Medical Air Systems Firmware	All	All	All	All
Operating System	Beaconmedaes	Scroll Medical Air Systems Firmware	All	All	All	All
cpe:2.3:h:beaconmedaes:scroll_medical_air_systems:-:*:*:*:*:*:						
cpe:2.3:h:beaconmedaes:scroll_medical_air_systems:-:*:*:*:*:*:						
cpe:2.3:o:beaconmedaes:scroll_medical_air_systems_firmware:*:*:*:*:*:						
cpe:2.3:o:beaconmedaes:scroll_medical_air_systems_firmware:*:*:*:*:*:						

No vendor comments have been submitted for this CVE

[← Previous ID](#)

Next ID→