




CVE-2020-36547

Published on: Not Yet Published

Last Modified on: 06/30/2022 06:07:11 PM UTC

CVE-2020-36547

[Source: Mitre](#)[Source: NIST](#)[CVE.ORG](#)[Print: PDF](#) 

Certain versions of [Voluson S8](#) from [Ge](#) contain the following vulnerability:

A vulnerability was found in GE Voluson S8. It has been rated as critical. This issue affects the Service Browser which itroduces hard-coded credentials. Attacking locally is a requirement. It is recommended to change the configuration settings.

CVE-2020-36547 has been assigned by [ics-cert@hq.dhs.gov](#) to track the vulnerability - currently rated as **HIGH** severity.

Affected Vendor/Software: [GE](#) - **Voluson S8** version n/a

CVSS3 Score: **7.8 - HIGH**

Attack Vector	Attack Complexity	Privileges Required	User Interaction
LOCAL	LOW	LOW	NONE
Scope	Confidentiality Impact	Integrity Impact	Availability Impact
UNCHANGED	HIGH	HIGH	HIGH

CVSS2 Score: **7.2 - HIGH**

Access Vector	Access Complexity	Authentication
LOCAL	LOW	NONE
Confidentiality Impact	Integrity Impact	Availability Impact
COMPLETE	COMPLETE	COMPLETE

CVE References

Description	Tags	Link
Veröffentlichung Schwachstellen in Ultraschallgeräten	www.scip.ch text/html	MISC www.scip.ch/?news.20200701

By selecting these links, you may be leaving CVEreport webspace. We have provided these links to other websites because they may have information that would be of interest to you. No inferences should be drawn on account of other sites being referenced, or not, from this page. There may be other websites that are more appropriate for your purpose. CVEreport does not necessarily endorse the views expressed, or concur with the facts presented on these sites. Further, CVEreport does not endorse any commercial products that may be mentioned on these sites. Please address comments about any linked pages to comment@cve.report.

There are currently no QIDs associated with this CVE

Known Affected Configurations (CPE V2.3)

Type	Vendor	Product	Version	Update	Edition	Language
Hardware	Ge	Voluson S8	-	All	All	All
Operating System	Ge	Voluson S8 Firmware	-	All	All	All
cpe:2.3:h:ge:voluson_s8:-:*:*:*:*:*:						
cpe:2.3:o:ge:voluson_s8_firmware:-:*:*:*:*:*:						

Discovery Credit

Marc Ruef/Rocco Gagliardi

Social Mentions

Source	Title	Posted (UTC)
@CVEreport	CVE-2020-36547 : A vulnerability was found in GE Voluson S8. It has been rated as critical. This issue affects the... twitter.com/i/web/status/1...	2022-06-17 13:16:02
@Robo_Alerts	Potentially Critical CVE Detected! CVE-2020-36547 A vulnerability was found in GE Voluson S8. It has been rated as... twitter.com/i/web/status/1...	2022-06-17 15:55:56
/r/netcve	CVE-2020-36547	2022-06-17 14:38:09

[← Previous ID](#)

[Next ID →](#)

© [CVE.report](#) 2023 |

Use of this information constitutes acceptance for use in an AS IS condition. There are NO warranties, implied or otherwise, with regard to this information or its use. Any use of this information is at the user's risk. It is the responsibility of user to evaluate the accuracy, completeness or usefulness of any information, opinion, advice or other content. EACH USER WILL BE SOLELY RESPONSIBLE FOR ANY consequences of his or her direct or indirect use of this web site. ALL WARRANTIES OF ANY KIND ARE EXPRESSLY DISCLAIMED. This site will NOT BE LIABLE FOR ANY DIRECT, INDIRECT or any other kind of loss.

CVE, CWE, and OVAL are registered trademarks of [The MITRE Corporation](#) and the authoritative source of CVE content is [MITRE's CVE web site](#). This site includes MITRE data granted under the following [license](#).

CVE.report and Source URL Uptime Status [status.cve.report](#)