



CVE-2021-23236

[MITRE](#)[NVD](#)[CVE.ORG](#)[JSON API](#)[Print: PDF](#)

Summary

CVE	CVE-2021-23236
State	PUBLIC
Assigner	ics-cert@hq.dhs.gov
Source Priority	CVE Program / NVD first with legacy fallback
Published	2022-01-21 19:15:00 UTC
Updated	2022-01-28 16:13:00 UTC
Description	Requests may be used to interrupt the normal operation of the device. When exploited, Fresenius Kabi Agilia Link+ version

Risk And Classification

Problem Types: CWE-400

NVD Known Affected Configurations (CPE 2.3)

Type	Vendor	Product	Version	Update	Edition	Language
Hardware	Fresenius-kabi	Agilia Connect	-	All	All	All
Operating System	Fresenius-kabi	Agilia Connect Firmware	All	All	All	All
Application	Fresenius-kabi	Agilia Partner Maintenance Software	All	All	All	All
Hardware	Fresenius-kabi	Link Agilia	-	All	All	All
Operating System	Fresenius-kabi	Link Agilia Firmware	All	All	All	All
Operating System	Fresenius-kabi	Link Agilia Firmware	3.0	-	All	All
Operating System	Fresenius-kabi	Link Agilia Firmware	3.0	d15	All	All
Application	Fresenius-kabi	Vigilant Centerium	1.0	All	All	All
Application	Fresenius-kabi	Vigilant Insight	1.0	All	All	All
Application	Fresenius-kabi	Vigilant Mastermed	1.0	All	All	All
Application	Fresenius-kabi	Agilia Connect	All	All	All	All
Application	Fresenius-kabi	Agilia Link	All	All	All	All
Application	Fresenius-kabi	Agilia Link	3.0	d15	All	All
Application	Fresenius-kabi	Agilia Partner Maintenance	All	All	All	All
Application	Fresenius-kabi	Vigilant Software Suite	1.0	All	All	All

References

Reference	Source	Link	Tags
Fresenius Kabi Agilia Connect Infusion System CISA	MISC	www.cisa.gov	
CVE Program record	CVE.ORG	www.cve.org	canonical
NVD vulnerability detail	NVD	nvd.nist.gov	canonical, analysis

Vendor Comments And Credit

Discovery Credit

LEGACY: Julian Suleder (ERNW Research GmbH), Nils Emmerich (ERNW Research GmbH), Raphael Pavlidis (ERNW Research GmbH), and Dr. Oliver Matula (ERNW Enno Rey Netzwerke GmbH) reported these vulnerabilities to the German Federal Office for Information Security (BSI) in the context of the BSI project ManiMed (Medical Device Manipulation Project).

There are currently no legacy QID mappings associated with this CVE.