



CVE-2021-27456

[MITRE](#)[NVD](#)[CVE.ORG](#)[Print: PDF](#)

Summary

CVE	CVE-2021-27456
State	PUBLIC
Assigner	ics-cert@hq.dhs.gov
Source Priority	CVE Program / NVD first with legacy fallback
Published	2022-03-23 20:15:00 UTC
Updated	2022-04-12 18:09:00 UTC
Description	Philips Gemini PET/CT family software stores sensitive information in a removable media device that does not have built-in

Risk And Classification

Problem Types: CWE-922

NVD Known Affected Configurations (CPE 2.3)

Type	Vendor	Product	Version	Update	Edition	Language
Hardware	Phillips	Gemini 882160	-	All	All	All
Operating System	Phillips	Gemini 882160 Firmware	-	All	All	All
Hardware	Phillips	Gemini 882300	-	All	All	All
Operating System	Phillips	Gemini 882300 Firmware	-	All	All	All
Hardware	Phillips	Gemini 882390	-	All	All	All
Operating System	Phillips	Gemini 882390 Firmware	-	All	All	All
Hardware	Phillips	Gemini 882400	-	All	All	All
Operating System	Phillips	Gemini 882400 Firmware	-	All	All	All
Hardware	Phillips	Gemini 882410	-	All	All	All
Operating System	Phillips	Gemini 882410 Firmware	-	All	All	All
Hardware	Phillips	Gemini 882412	-	All	All	All
Operating System	Phillips	Gemini 882412 Firmware	-	All	All	All
Hardware	Phillips	Gemini 882470	-	All	All	All
Operating System	Phillips	Gemini 882470 Firmware	-	All	All	All
Hardware	Phillips	Gemini 882471	-	All	All	All
Operating System	Phillips	Gemini 882471 Firmware	-	All	All	All
Hardware	Phillips	Gemini 882473	-	All	All	All

Operating System	Phillips	Gemini 882473 Firmware	-	All	All	All
Hardware	Phillips	Gemini 882476	-	All	All	All
Operating System	Phillips	Gemini 882476 Firmware	-	All	All	All
Hardware	Phillips	Truflight 882438	-	All	All	All
Operating System	Phillips	Truflight 882438 Firmware	-	All	All	All

References

Reference	Source	Link	Tags
Philips Gemini PET/CT Family CISA	CONFIRM	www.cisa.gov	
Product Security Philips	CONFIRM	www.philips.com	
CVE Program record	CVE.ORG	www.cve.org	canonical
NVD vulnerability detail	NVD	nvd.nist.gov	canonical, analysis

Vendor Comments And Credit

Discovery Credit

LEGACY: Jean GEORGE – CHU UCL Namur – Nuclear medicine department reported this vulnerability to Philips.

There are currently no legacy QID mappings associated with this CVE.

© CVE.report 2026 |

Use of this information constitutes acceptance for use in an AS IS condition. There are NO warranties, implied or otherwise, with regard to this information or its use. Any use of this information is at the user's risk. It is the responsibility of user to evaluate the accuracy, completeness or usefulness of any information, opinion, advice or other content. EACH USER WILL BE SOLELY RESPONSIBLE FOR ANY consequences of his or her direct or indirect use of this web site. ALL WARRANTIES OF ANY KIND ARE EXPRESSLY DISCLAIMED. This site will NOT BE LIABLE FOR ANY DIRECT, INDIRECT or any other kind of loss.

CVE, CWE, and OVAL are registered trademarks of [The MITRE Corporation](https://www.mitre.org/) and the authoritative source of CVE content is [MITRE's CVE web site](https://www.mitre.org/cve/). This site includes MITRE data granted under the following [license](https://www.mitre.org/licenses/).

CVE.report and Source URL Uptime Status status.cve.report