



CVE-2021-33886

[MITRE](#)[NVD](#)[CVE.ORG](#)[JSON API](#)[Print: PDF](#)

Summary

CVE	CVE-2021-33886
State	PUBLIC
Assigner	cve@mitre.org
Source Priority	CVE Program / NVD first with legacy fallback
Published	2021-08-25 12:15:00 UTC
Updated	2022-07-12 17:42:00 UTC
Description	An improper sanitization of input vulnerability in B. Braun SpaceCom2 prior to 012U000062 allows a remote unauthenticated

Risk And Classification

Problem Types: CWE-134

NVD Known Affected Configurations (CPE 2.3)

Type	Vendor	Product	Version	Update	Edition	Language
Hardware	Bbraun	Infusomat Large Volume Pump 871305u	-	All	All	All
Operating System	Bbraun	Spacecom2	All	All	All	All
Hardware	Bbraun	Spacestation 8713142u	-	All	All	All

References

Reference	Source	Link
404 Not Found	MISC	www.bbraunusa.com
McAfee Enterprise ATR Uncovers Vulnerabilities in Globally Used B. Braun Infusion Pump McAfee Blogs	MISC	www.mcafee.com
CVE Program record	CVE.ORG	www.cve.org
NVD vulnerability detail	NVD	nvd.nist.gov

No vendor comments have been submitted for this CVE.

There are currently no legacy QID mappings associated with this CVE.

this information or its use. Any use of this information is at the user's risk. It is the responsibility of user to evaluate the accuracy, completeness or usefulness of any information, opinion, advice or other content. EACH USER WILL BE SOLELY RESPONSIBLE FOR ANY consequences of his or her direct or indirect use of this web site. ALL WARRANTIES OF ANY KIND ARE EXPRESSLY DISCLAIMED. This site will NOT BE LIABLE FOR ANY DIRECT, INDIRECT or any other kind of loss.

CVE, CWE, and OVAL are registered trademarks of [The MITRE Corporation](#) and the authoritative source of CVE content is [MITRE's CVE web site](#). This site includes MITRE data granted under the following [license](#).

Free CVE JSON API [cve.report/api](#)

CVE.report and Source URL Uptime Status [status.cve.report](#)