



# CVE-2023-30561

[MITRE](#)[NVD](#)[CVE.ORG](#)[Print: PDF](#)

## Summary

<b>CVE</b>	CVE-2023-30561
<b>State</b>	PUBLIC
<b>Assigner</b>	cybersecurity@bd.com
<b>Source Priority</b>	CVE Program / NVD first with legacy fallback
<b>Published</b>	2023-07-13 20:15:00 UTC
<b>Updated</b>	2023-07-25 18:51:00 UTC
<b>Description</b>	The data flowing between the PCU and its modules is insecure. A threat actor with physical access could potentially read or

## Risk And Classification

**Problem Types:** CWE-311

## NVD Known Affected Configurations (CPE 2.3)

Type	Vendor	Product	Version	Update	Edition	Language
Hardware	Bd	Alaris 8015 Pcu	-	All	All	All
Operating System	Bd	Alaris 8015 Pcu Firmware	All	All	All	All

## References

Reference	Source	Link	Tags
BD Alaris™ System with Guardrails™ Suite MX	MISC	<a href="http://www.bd.com">www.bd.com</a>	
CVE Program record	CVE.ORG	<a href="http://www.cve.org">www.cve.org</a>	canonical
NVD vulnerability detail	NVD	<a href="http://nvd.nist.gov">nvd.nist.gov</a>	canonical, analysis

No vendor comments have been submitted for this CVE.

There are currently no legacy QID mappings associated with this CVE.

CVE, CWE, and OVAL are registered trademarks of [The MITRE Corporation](#) and the authoritative source of CVE content is [MITRE's CVE web site](#). This site includes MITRE data granted under the following [license](#).

**CVE.report and Source URL Uptime Status [status.cve.report](#)**